

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the Measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer **MeterSIt**  
Viale dell'Industria, 31  
35129 Padova  
Italy

Measuring instrument **A thermal-mass flow gas meter**  
Type : x4850xx  
x4851xx  
x4852xx  
x4853xx  
x4854xx  
x4855xx

Manufacturer's mark or name : MeterSIt

Destined for the measurement of : Gas volume of natural gas or  
hydrogen.  
See description §1.2.3

Accuracy class : Class 1,5

Environment classes : M1 / E2

Gas temperature range : -25 °C / +55 °C

Ambient temperature range : -25 °C / +55 °C

Designed for : Condensing humidity

Intended location : Open

Further properties are described in the annexes:

- Description T10362 revision 55;
- Documentation folder T10362-41.

Valid until 1 July 2031

Remark This revision replaces the previous versions, including its documentation  
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
3 February 2023

## Certification Board

**NMI Certin B.V.**  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that  
no liability is accepted and that the  
manufacturer shall indemnify third-party  
liability.

The designation of NMI Certin B.V. as Notified  
Body can be verified at [http://  
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete  
document only is permitted.

This document is digitally signed  
and sealed. The digital signature  
can be verified in the blue ribbon at  
the top of the electronic version of  
this certificate.